

HEV Seroconversion Panel

REF Catalogue No: SCP-HEV-007a

The Biomex GmbH HEV Seroconversion Panel consists of 17 members with each member containing 1.0 mL of human plasma. This panel illustrates the onset and decline of IgM and IgG Hepatitis E virus antibodies from one individual over a period of 152 days.

1. Intended Use

This Seroconversion Panel (SCP) is intended for standard testing by diagnostic manufacturers and researchers during assay development, evaluation, troubleshooting and post-marked surveillance of IgM/IgG antibody test systems and methods. Moreover, it serves as validation tool for diagnostic sensitivity, determination of analytical sensitivity, identification of cut-off values or to study the humoral immune response to this infection.

2. Storage and Stability

Store the SCP at -20°C to -80°C. Thaw samples at room temperature and mix gently by inversion before usage. Avoid foaming, contamination and repeated freeze and thaw cycles. After usage, return immediately to storage conditions.

3. Warnings and Precautions

Potentially infectious materials. Handle the product as if capable of transmitting infectious diseases. Do not pipette by mouth.

Prevention:

P264: Wash hands thoroughly after handling.

P270: Do not eat, drink or smoke when using this product

P273: Avoid the release to the environment

Disposal:

P501: Dispose of waste in accordance to applicable local or national regulations. Waste must be disposed in a secured manner.

4. Donor Information

All panel members have been tested and found negative/non-reactive for anti-HIV 1/2, anti-HCV and HBsAg with CE marked tests.

Donor profile

- Sex: Male
- Age: 26
- Ethnicity: Caucasian
- Residence: Germany
- HEV Genotype: 3c

5. Detection Methods

Each panel member has been tested for HEV RNA with RealStar HEV RT-PCR kit by Altona Diagnostic Technologies (ADT). The HEV IgM and IgG antibodies were detected with Mikrogen recomWell. Testing is performed with CE marked test.

6. Limitations and Restrictions

This panel is for Research Use Only and not intended for human or animal diagnostics, or for therapeutic purposes. Each laboratory has the responsibility to ascertain the suitability of the SCP for its particular application and to establish their own guidelines for interpretation of results. Data is provided for informational purposes only. The Biomex GmbH does not claim that others can duplicate these test results exactly.

Declaration of used symbols

REF	LOT	i	 Manufactured by	2°C - 8°C Temperature Limitation
Catalog Number	Lot Number	Consult Instructions for Use		

HEV Seroconversion Panel, Catalogue No: SCP-HEV-007a

Panel Member	Day	Blood Date	RealStar HEV RT PCR IU/mL	Mikrogen recom-Well HEV IgM U/mL	Mikrogen recomWell HEV IgM Valuation	Mikrogen recomWell HEV IgG U/mL	Mikrogen recomWell HEV IgG Valuation	GLDH U/L	ALT U/L	AST U/L	BiliT mg/dL
				cut off 24 U/mL		cut off 24 U/mL					
1	0	19.08.2011	12.8	6.46	negative	3.20	negative	3.7	26	33	0.23
2	4	23.08.2011	246	3.90	negative	2.78	negative	5.2	30	29	0.16
3	7	26.08.2011	347	3.04	negative	2.68	negative	4.9	18	29	0.15
4	19	07.09.2011	7310	5.27	negative	3.19	negative	4.4	25	27	0.20
5	24	12.09.2011	17700	11.69	negative	2.86	negative	3.5	23	25	0.47
6	28	16.09.2011	219000	5.83	negative	3.20	negative	11.1	91	69	0.34
7	32	20.09.2011	21800	12.42	negative	8.62	negative	11.8	109	34	0.30
8	39	27.09.2011	1920	9.54	negative	25.95	positive	10.1	44	35	0.21
9	46	04.10.2011	43.8	8.30	negative	43.83	positive	6.1	24	36	0.14
10	52	10.10.2011	17.5	6.32	negative	47.74	positive	3.8	24	28	0.44
11	59	17.10.2011	negative	8.74	negative	43.27	positive	5.3	26	30	0.40
12	62	20.10.2011	negative	9.41	negative	44.27	positive	3.2	18	24	0.28
13	66	24.10.2011	negative	5.17	negative	41.27	positive	4.5	21	28	0.40
14	69	27.10.2011	negative	6.19	negative	39.01	positive	2.5	19	26	0.33
15	103	30.11.2011	negative	3.78	negative	40.58	positive	3.2	19	26	0.21
16	145	11.01.2012	NT	NT	NT	NT	NT	NT	NT	NT	NT
17	152	18.01.2012	NT	NT	NT	NT	NT	NT	NT	NT	NT

NT= not tested

HEV Seroconversion Panel, Catalogue No: SCP-HEV-007a

